

# FAQ

Subject : How to use multi probes(MG07.1 + External Probe)

Date : 2020 06 11

## <Example>



**GUVV-T10GS7.1-LA9**

**100mW/cm<sup>2</sup> Class**

(0~99.99mW/cm<sup>2</sup> measuring)



**Available**

UVB : **100mW/cm<sup>2</sup> Class**  
(0~99.99mW/cm<sup>2</sup> measuring )



**Available**

UVC : **100mW/cm<sup>2</sup> Class**  
(0~99.99mW/cm<sup>2</sup> measuring )



**Not available**

Visible : **10mW/cm<sup>2</sup> Class**  
(0~9.999mW/cm<sup>2</sup> measuring )

**When the measuring range of the probe is the same  
It is possible to replace multiple probes in one monitor.**

**\*Integral products is not possible**