


# FAQ

Subject : How to adjust the output of LA6 probe gain

Date : 2020 05. 12



Fig1. LA6  
Probe



Turning the arrow counterclockwise increase the output value.

Turning the arrow clockwise decrease the output value.

- \* Shipment calibration setting changed after gain adjustment.
- \* If you want the same value as the shipment cal. value, you must re-calibration.  
(Please note that you should pay the charge for both re-calibration and freight)